

12-1-2019

Section: Mathematics, Statistics, Computer Science, Physics and Astronomy

RELIABILITY OF REUSING THE SENSITIZED LIF THERMOLUMINESCENCE DOSIMETERS

M. Farag

Ionizing Radiation Metrology Department, National Institute for Standards, El-Haram, P.O. 136, Code No 12211, Giza, Egypt., maha_6_86@yahoo.com

A. Sadek

Ionizing Radiation Metrology Department, National Institute for Standards, El-Haram, P.O. 136, Code No 12211, Giza, Egypt

H. Shousha

Ionizing Radiation Metrology Department, National Institute for Standards, El-Haram, P.O. 136, Code No 12211, Giza, Egypt.

A. Al-Hagg

Physics Department, Faculty of Science, Al-Azhar University of Girl's Branch, 11884, Cairo, Egypt.

G. Kitis

Aristotle University of Thessaloniki, Nuclear Physics Laboratory, 54124, Thessaloniki, Greece

Follow this and additional works at: <https://absb.researchcommons.org/journal>



Part of the [Life Sciences Commons](#)

How to Cite This Article

Farag, M.; Sadek, A.; Shousha, H.; Al-Hagg, A.; and Kitis, G. (2019) "RELIABILITY OF REUSING THE SENSITIZED LIF THERMOLUMINESCENCE DOSIMETERS," *Al-Azhar Bulletin of Science*: Vol. 30: Iss. 2, Article 2.

DOI: <https://doi.org/10.21608/absb.2019.86750>

This Original Article is brought to you for free and open access by Al-Azhar Bulletin of Science. It has been accepted for inclusion in Al-Azhar Bulletin of Science by an authorized editor of Al-Azhar Bulletin of Science. For more information, please contact kh_Mekheimer@azhar.edu.eg.

RELIABILITY OF REUSING THE SENSITIZED LiF THERMOLUMINESCENCE DOSIMETERS

Farag; M. A.^{1*}, Sadek; A. M.¹, Shousha; Hany. A.¹, Al-Hagg; A. A.², Kitis; G.³.

1. Ionizing Radiation Metrology Department, National Institute for Standards, El-Haram, P.O. 136, Code No 12211, Giza, Egypt.

2. Physics Department, Faculty of Science, Al-Azhar University of Girl's Branch, 11884, Cairo, Egypt.

3. Aristotle University of Thessaloniki, Nuclear Physics Laboratory, 54124, Thessaloniki, Greece

Corresponding Author : maha_6_86@yahoo.com

ABSTRACT

The reliability of reusing the sensitized thermoluminescence (TL) for LiF (TLD-100) dosimeters have been investigated. It has been found that reusing the same sensitized LiF dosimeters may influence the sensitivity of the dosimeters even under the same experimental conditions when the temperature readout range does not include the high-temperature TL signal at $T > 505\text{K}$. But, if the high-temperature TL signal is included in the temperature readout range, it has been found that the ultra-high temperature TL signal significantly increases with increasing the number of reusing the dosimeters. After almost 7 times of reusing the sensitized dosimeters, the total area under the glow-curve becomes stable and independent of the number of reusing the dosimeters. It has been also found that the TL signal located between the low-temperature TL signal and the ultra-high-temperature TL signal was almost stable over all the number of reusing the dosimeters. For all the low-temperature peaks, the kinetics parameters of the sensitized LiF dosimeters have not to be influenced by reusing the same dosimeters. However, the activation energy of peak 5 which lies at temperature range (460-500K) has been influenced by the readout temperature range and reusing the dosimeters. The typical value for this glow-peak has been obtained in the case which the high-temperature TL signal has been included in the temperature readout.

Keywords: Sensitization; Glow curve analysis; Successive irradiation; Readout temperature.

1. INTRODUCTION

Thermoluminescence dosimetry (TLD) is a well-established dosimetric technique with applications in areas such as personnel, environmental and clinical dosimetry. Therefore, considerable research was carried out for a better understanding and improvement of the material characteristics as well as to develop new materials [1]. Nowadays, there is an intensive development of radiation technologies employing high-dose irradiation to modify material properties [2]. It was found that the sensitivity of the TL dosimeters subjected to high dose irradiation followed by the thermal treatment could be enhanced [3]. An often-quoted example is provided by LiF: Mg, Ti in which the sensitivity of TL peak 5 which lies at temperature range (460-500K) could be increased by giving the sample a large pre-dose (approximately 10 Gy) followed by an anneal at 300°C [3, 4]. Different sensitization studies were also investigated to enhance the sensitivity of the LiF detectors (e.g., [5, 6, 7, 8, 9, 10]).

Farag et al. [11] have developed sensitization procedure for LiF (TLD-700) dosimeters. It has been found the post-irradiation anneal at 553K (280°C) for 30 min and a pre-test dose of 2 kGy have increased the sensitivity (total area under the glow-curve) of the dosimeters by a factor of ~35. However, the dominant factor was increasing the sensitivity of the high-temperature peak (peak-7) by a factor of ~22. The main idea of the sensitization procedure is to reduce the effect of the deep traps competitors via irradiation and thermal treatments [4]. By suitable pre-irradiation and thermal treatment, the active traps are left empty while the deep trap is left full. In this way, the competitor is removed and subsequent irradiation of the sample leads to an enhanced TL signal [12].

In the routine thermoluminescence (TL) dosimetry, the same samples are being reused several times. Therefore, the samples shall be subject to thermal annealing to be reset and hence be reused. However, it was reported that the repeated use of the TL sample may change the sensitivity over the usage period. This is because the repeated use repeatedly places charge into the deep traps and this charge accumulates with each reuse [4].

Another important factor that may influence the sensitivity of the dosimeters is the repeated thermal treatments. In other words, the high-temperature thermal treatments may evacuate the deep traps of the sensitized dosimeters and therefore affects the sensitivity of the dosimeters [13]. The high-temperature thermal treatment could be due to the thermal annealing or the temperature during the readout process. In the work of Farag *et al.* [11], the thermal annealing for the sensitized LiF dosimeters has been adjusted to 553K (280°C) which cannot be considered high temperature.

The main aim of the present work is to investigate the stability of the sensitization procedure developed by Farag *et al.* [11] via two steps:

- i. Investigate the effect of the temperature readout range, and
- ii. Investigate the effect of the repeated reuse of the same dosimeters on the sensitivity as well as the kinetics parameters of the sensitized LiF dosimeters.

2. EXPERIMENTAL WORK

2.1. Material and methods

In the current study, TLD-100 dosimeters of size $(3.2 \times 3.2 \times 0.9)$ mm³ have been used. A separate calibrated one-cycle electric oven manufactured by Fisher Scientific Company with an accuracy of $\pm 1^\circ\text{C}$ has been used for the thermal treatments. The irradiation processes have been performed using a Co-60 gamma source delivering a dose rate of 1.23 Gy min^{-1} . The readout process has been conducted using Harshaw-3500 manual reader with a linear rate of 2°C s^{-1} with different readout temperature zones.

2.2. Investigate the effect of the temperature readout range

In order to investigate the influence of the temperature readout range on the stability of the sensitized LiF dosimeters, the following experimental procedures have been followed

1. The TLD-100 dosimeters have been subjected to the sensitization procedure [11].
2. The sensitized detectors have been subjected to a 0.5 Gy test dose.
3. The dosimeters have been divided into five groups G1 – G5. The temperature readout range of each group is described in Table 1.

Table 1. The temperature readout range applied to the sensitized TL dosimeters.

Group Number	Readout Temperature Zone (K)
G1	323 – 553
G2	323 – 573
G3	323 – 593
G4	323 – 613
G5	323 – 633

2.3 Investigate the effect of the repeated use

The above experimental procedures have been repeated 10 times using the same dosimeters for each readout temperature range (presented in Table 1) to investigate the effect of the repeated use.

3. GLOW-CURVE DECONVOLUTION ANALYSIS

In the current study, there was a need to calculate the area under certain components of the glow curves. Indeed, the peak height determined from the experimental glow-curve can be also used. However, this method, especially in the present study, may not be appropriate due to the complexity of the glow-curves (as presented in Fig.5). In addition to the high fluctuation in the results reported by this method [14]. The alternative method is using the routine glow-curve deconvolution analysis by fitting the glow-curve with a mathematical TL model. In the current study, the general-order kinetics (GOK) model has been employed in the deconvolution analysis process. The mathematical representation of this model is given by [15];

$$I(T) = I_m (b)^{\frac{b}{b-1}} \exp\left(\frac{E}{kT} \frac{T - T_m}{T_m}\right) \left[(b-1)(1-\Delta) \frac{T^2}{T_m^2} \exp\left(\frac{E}{kT} \frac{T - T_m}{T_m}\right) + Z_m \right]^{-\frac{b}{b-1}} \quad (1)$$

Where I_m and T_m (K) are the peak maximum and peak maximum position, respectively, E (eV) is the activation energy, b is the order of kinetics which changes from ~ 1 to 2 as the glow-peak changes from the first- to the second-order kinetics with $\Delta = 2kT/E$, $\Delta_m = 2kT_m/E$ and $Z_m = 1 + (b-1) \Delta_m$. The advantage of using this equation is that this expression uses the peak maximum and the peak maximum position in which their initial values can be determined from the experimental glow-curve.

The fitting process has been performed using the ‘trust-region’ optimization algorithm [16] via the MATLAB package. This algorithm is a simple yet powerful concept optimization. It can solve difficult non-linear problems more efficiently than the other algorithms and it represents an improvement over the popular Levenberg-Marquardt algorithm [17]. In this algorithm, the ‘starting value’ and/or lower and upper ‘limits’ of the unknown parameters can be set. This algorithm was used in the TL glow-curve analysis for different applications [18, 19, 20, 21, 22, 23] including the analysis of complicated and overlapping glow-peaks. The quality of the fitting process has been assessed using the Figure of Merit (FOM) in the work of Balian and Eddy [24]

$$\text{FOM}(\%) = \sum_{j_i}^{j_f} \frac{|y_i - y(x_i)|}{A} \times 100 \quad (2)$$

Where FOM is the figure of merit, j_i is the first channel in the region of interest, j_f is the last channel in the region of interest, y_i is the information content of channel j , $y(x_i)$ is the value of the fitting function in channel j , and A is the integral of the fitted glow-peak in the region of interest. In the current work, most of FOM values did not exceed 2%.

Nevertheless, in some complicated cases, where the deconvolution analysis is meaningless, there is a simple integration of the TL signal between two predetermined temperatures has been used to compute the area under the desired TL signal region. This method has been used, especially, to compute the area under the high-temperature TL region because of the complexity of this region and non-universal agreement on the deconvolution procedures of this region [25, 23].

4. RESULTS AND DISCUSSION

4.1. Influence of temperature readout range on the glow-curve of the sensitized LiF dosimeters

In the current study, the dosimeters have been subjected to the same experimental conditions but with different temperature readout ranges (Table 1). The glow-curves of the first and last groups G1

and G5 are presented in Fig.1. The typical glow-curve of LiF dosimeter in case of gamma irradiation consists of four low-temperature glow-peaks (2-5) with the main dosimetric peak 5 which is the dominant TL signal in the dosimetric applications [26]. The high-temperature TL signal (HTTL) in LiF dosimeters is the TL signal located in the temperature region of [$\sim 505 - \sim 550$]K as illustrated in Fig.1. The analysis of this TL signal was discussed in the work of Horowitz and Yossian [25]. Unusual ultra-HTTL signal has been observed for the glow-curves of sensitized LiF dosimeters recorded with temperature readout range of [323-633]K (Fig.1). This signal has been reported in the work of Farag *et al.* [11] and was attributed to the effect of the sensitization process subjected to the LiF dosimeters. The area under the low-temperature peaks and the HTTL signal have been calculated using the curve-integral method. The area under the ultra-HTTL could not be included because this signal has been observed only at the [323-633]K temperature readout range.




Fig. 1. Effect of different temperature readout ranges on sensitized TLD-100 dosimeters.




Fig. 2. The area under the low-temperature peaks and the HTTL as a function of the readout temperature zone.

The glow-curves presented in Fig.1 and the analysis results in Fig.2 imply that changing the temperature readout range has no significant changes neither on the sensitivity nor on the shape of the glow-curves. Actually, this was expected because there is no strong interaction between the low-temperature and the high-temperature peaks during the thermal stage. Of course, the interactions have been already taken place during the “irradiation stage” in which the capacity and the trapping probabilities are the dominant factors.

4.2. Reliability of reusing the sensitized LiF dosimeters

4.2.1. Effect on the TL glow-curve

In the TL routine dosimetry work, the same dosimeters are used many times. For this reason, the dosimeters should be subjected to certain thermal annealing conditions to reset the dosimeters. It is believed that after the thermal annealing process, all the information due to the previous irradiation process were erased. However, it was reported that reusing the same dosimeters several times may have influenced the sensitivity of the dosimeters even with applying the recommended thermal annealing treatment [27].

In the work of Farag *et al.* [11], the developed sensitization procedure is mainly based on reducing the competition effect of the deep traps on the active traps. This could be achieved via two steps; the first step is to fill these deep traps with electrons until approaching the saturation level. While the second step is to use the appropriate thermal treatments so that the trapped electrons in the deep traps cannot be released. However, the influence of reusing the dosimeters several times on their sensitivity has not been addressed. In the current section, the effect of reusing the sensitized LiF dosimeters has been investigated throughout the different ranges of the temperature readout. The same

experiment carried out in section 4.1 has been repeated 10 times for each temperature readout. As examples, the glow-curves in cases of the [323 – 553]K and [323 – 633]K temperature readout ranges are illustrated in Fig.3.




Fig. 3. An example of the glow curves on the influence of re-use the sensitized dosimeters for run-1 and run-10 in cases of the [323 – 553]K and [323 – 633]K temperature readout ranges.

Regardless of the change in the sensitivity of the TL signal, the shape of the glow-curve and the position of the glow-peak components have not been changed. However, in case of the [323 – 633]K temperature readout range, the ultra-HTTL signal has become the dominant signal in the glow-curve.

4.2.2. Effect on the TL sensitivity

It is obvious from Fig.3 that reusing the same dosimeters several times have caused a dramatic effect on the sensitivity of the dosimeters. The main dosimetric peak 5 component has significantly decreased as the number of reusing the dosimeters increased. Fig.4 presents the total area under the glow-curves recorded as the number of reusing the dosimeters increase.




Fig. 4. The sensitivity of sensitized LiF dosimeters as a function of the number of reusing the dosimeters.

After reusing the dosimeters 10 times, the sensitivity of the dosimeters was decreased by a factor ~ 0.6 in all the cases, except in the case of [323 – 633]K temperature readout range. In this case, after

reusing the dosimeters 10 times, the sensitivity of the dosimeters remained almost the same. However, an abrupt change has been observed in the sensitivity of the dosimeters when repeating using the dosimeters 4 – 7 times. To investigate this behavior, the glow-curves have been deconvoluted as illustrated in Fig.5.




Fig. 5. An example of the deconvolution process conducted to the glow-curves of sensitized LiF dosimeters for reusing run-10 in case of the [323 – 633]K temperature readout range.

The low-temperature TL signal has been deconvoluted into the usual four low-temperature peaks (2-5). The HTTL signal has been deconvoluted into 5 components to approach the best goodness-of-fit. In fact, there is no unique solution to the analysis of the HTTL. Bos *et al.* [28] and Perks and Marshall [29] deconvoluted the HTTL into six and seven components, respectively. While Horowitz *et al.* [30] and Horowitz *et al.* [31] deconvoluted the HTTL into seven peaks. However, Sadek *et al.* [23] have deconvoluted the HTTL of LiF dosimeters irradiated with Alpha particles into 3 glow-peaks. Indeed, the aim in the current section is neither to address the deconvolution solution of this portion nor to evaluate the kinetics parameters of its component but rather to compute the sensitivity (*i.e.* the area under) this portion and separate it from the ultra-HTTL signal. Therefore, the area under this HTTL signal has been computed as the summation of the area under the 5 deconvoluted glow-peak. Moreover, an exponential function has been used to compute the area under the ultra-HTTL signal. The results are presented in Fig.6.




Fig. 6. The sensitivity of low-temperature peaks, high-temperature TL signal HTTL, and ultra-high temperature TL (Ultra-HTTL) in case of the [323- 633] K temperature readout range as a function of repeatability number.

It is obvious from Fig.6a that the area under peak 5 is gradually decreasing with increasing the number of reusing the dosimeters. This observation has been also reported for the glow-peaks 2, 3 and 4 (Fig.6b). While a significant increase in the area under the ultra-HTTL portion has been observed as the number of reusing the dosimeters increased. This explains the increase in the total area observed in Fig.5 and it also implies that the area under this portion is dominant. However, after reusing the dosimeters 7 times, the sensitivity of the ultra-HTTL portion becomes almost stable.

It was proposed that the glow-peak 10 located at $\sim 673\text{K}$ works as a competitor for the glow-peak 5 [4]. The maximum of the temperature readout range in the current study has been set to 633 K . The successful reusing the dosimeters with temperature readout approaching this degree would evacuate some of the electrons trapped in this competitor, and hence, increasing the competition effect on peak 5. This may explain the decreasing in the TL signal of this peak and increasing the TL signal of the ultra-HTTL signal. After a certain reusing number, the rate of decreasing the TL signal of the low-temperature peaks is equivalent to the rate of increasing the ultra-HTTL signal, and thereby, the total area under the whole glow-curve appears stable.

It is worthwhile to mention that the area under HTTL was almost stable over increasing the reusing number. This may refer to the importance of this signal in the field of TL dosimetry. However, the complex overlapping of this signal with the neighbor signals and the non-unique deconvolution solution for this signal always restrict the using of this signal in the TL dosimetry. Therefore, one may state that the main conclusion of this section is that the total area under the glow-curve of sensitized dosimeters can be used as a dosimetric signal after reusing the dosimeters at least 7 times to approach the stability level.

4.3 Effect on the kinetic parameters

The influence of reusing the same dosimeters on the kinetics parameters has been addressed in the present section. The peak position and the activation energy values obtained for the glow-curves of sensitized LiF dosimeters readout using temperature range of [323-633] K for 10 times are presented in Fig.7.




Fig. 7. Peak Position and activation energy values for the low-temperature glow-peaks of sensitized LiF dosimeters readout under the same temperature readout range [323-633] K for 10 times.

The kinetics parameters of the low-temperature peaks have not been changed with increasing the reusing number. This was expected since there was no change in the shape or the position of the glow-peaks and therefore, the kinetics parameters have not been changed. However, the high activation energy for peak 5 has been obtained. The typical range of the activation energy values of peak 5 is [2.02-2.17] eV [28]. However, the values in the range [2.0-2.3] eV were considered high activation energy values for peak 5 [32]. Indeed, the high activation energy value observed for peak 5 have been reported by Farag *et al.* [11] and were attributed to the sensitization process. In other words, this high value is not due to reusing the dosimeters but rather due to the sensitization process. It is to be noted

that this observation has been also reported for all the readout temperature ranges performed in the current study.

The activation energy and the other kinetics parameters for the low-temperature glow-peaks have not been influenced by changing the temperature readout range (Fig.8). This was expected because it has been already discussed that including the high-temperature TL signal in the temperature readout range has not influenced the low-temperature glow-peaks.




Fig. 8. The activation energy values of the low-temperature peaks throughout the temperature readout ranges for reusing run-10.

The kinetics parameters reported for the low-temperature peaks are the typical values reported in the literature [28, 25, 33]. It implies that the competitor deep traps may influence the sensitivity of the glow-peaks but not their kinetics parameters. This conclusion has been recently reported by Sadek and Kitis [34] using theoretical simulation via interactive phenological TL models.

CONCLUSIONS

- In the case of [323 – 553] K temperature readout range, the main dosimetric peak 5 has become the dominant TL signal in the glow-curve, while in case of the [323 – 633]K temperature readout range, the ultra-HTTL signal has become the dominant TL signal in the glow-curve.
- After reusing the dosimeters 10 times, the sensitivity in the case of the total area of the dosimeters was decreased by a factor ~ 0.6 in all the cases of different temperature readout, except in the case of [323 – 633]K temperature readout range remained almost the same.
- The low-temperature signals decrease while the ultra-high-temperature signals increase with reusing the dosimeters several times.
- After reusing the sensitized dosimeters at least 7 times, the total area under the glow-curve becomes almost stable and reliable for use in the TL dosimetry.
- Therefore, adjusting the temperature readout range approaching this degree (~ 633) K to include the high-temperature TL signals is recommended for the sensitized LiF dosimeters.
- The high-temperature TL signal located at the temperature range of [323 – 553] K, namely peaks 7 and 8, are stable over reusing the same dosimeters several times.

- The kinetics parameters, the peak position, and the shape of the glow curve for the low-temperature peaks have neither been influenced by changing the temperature readout range nor reusing the dosimeters several times.
- High activation energy values for peak 5 have been observed. These high values are neither because of reusing the dosimeters several times nor changing the temperature readout range but rather because of the sensitization process.

6. REFERENCES

- [1] Bos A J J. High sensitivity thermoluminescence dosimetry. Nucl. Instrum and Meth in Phys. Res. B. 2001; (184): 3-28.
- [2] Kortov V, Ustyantsev Yu. Advantages and challenges of high-dose thermoluminescent detectors. Radiat. Meas. 2013; (56): 299-302.
- [3] Hwang F S W. Supralinearity and sensitization of thermoluminescent phosphor lithium fluoride. J. Phys. D: Appl. Phys. 1971; (4):598-607.
- [4] Chen R, McKeever S W S. Theory of thermoluminescence and related phenomena. 1997, World Scientific Publishing: Singapore.
- [5] Marrone M J, Attix F H. Damage effect in CaF₂: Mn and LiF thermoluminescent dosimeters. In Health Physics, Northern Ireland, Health Physics Press. 1964; (10):431-436.
- [6] Cameron J R, De Werd L A, Wanger J, Wilson C, Doppke K, Zimmerman D W. Symp. on solid-state and chemical radiation dosimetry in medicine and biology. 1966; SM-78/21: (Vienna, IAEA).
- [7] Jain V K, Kathuria S P, Ganguly A K. Radiation damage in thermoluminescent LiF TLD-phosphor. J. Phys. C: Solid State Phys. 1975; (8): 2191-2197.
- [8] Bhatt B C, Lakshmanan A R, Shinde S S, Bhatt R C. Radiation damage in LiF: Mg, Ti and CaSO₄: Dy TLD phosphors. Radiat. Prot. Dosim. 1988; 25(1): 37-41.
- [9] Cai G, Wang S, Zhang J. Radiation damage and recovery of LiF: Mg, Ti TLD at the dose of 10⁶ Gy γ -rays. Phys. Stat. Sol. (a). 1990; (121): 57-61.
- [10] Bilski P, Obryk B, Stuglik Z. Behavior of LiF: Mg, Cu, P and LiF: Mg, Ti thermoluminescent detectors for electron doses up to 1 MGy. Radiat. Meas. 2010; (45): 576-578.
- [11] Farag M A, Sadek A M, Shousha Hany A, El-Hagg A A, Atta M R, Kitis G. Radiation damage and sensitization effects on thermoluminescence of LiF: Mg, Ti (TLD-700). Nucl. Instrum and Meth in Phys. Res. B. 2017; (407): 180-190.
- [12] McKeever S W S. 5.5 eV optical absorption, supralinearity, and sensitization of thermoluminescence in LiF TLD-100. J. Appl. Phys. 1990; 68(2): 724-731.
- [13] Horowitz Y S. LiF: Mg, Ti versus LiF: Mg, Cu, P: The competition heats up. Radiat. Prot. Dosim. 1993; (47): 135-141.
- [14] Horowitz Y S, Moscovitch M. Highlights and pitfalls of 20 years of application of computerized glow curve analysis to thermoluminescence research and dosimetry. Radiat. Prot. Dosim. 2012; 153(1): 1-22.
- [15] Kitis G, Gómez-Ros J M, Tuyn J W N. Thermoluminescence glow-curve deconvolution functions for first, second and general orders of kinetics. J. Phys. D: Appl. Phys. 1998; (31): 2636-2641.
- [16] Moré J J, Sorensen D C. Computing a trust region step. SIAM J. Sci. Stat. Comput. 1983; 4 (3): 553-572.
- [17] Mathworks, MATLAB Curve Fitting Toolbox: User's Guide, USA: The MathWorks, Inc., 2004.
- [18] Sadek A M. Test of the accuracy of the computerized glow curve deconvolution algorithm for the analysis of thermoluminescence glow curves. Nucl. Instrum and Meth in Phys. Res. A. 2013; (712): 56-61.

- [19] Sadek A M, Eissa H M, Basha A M, Kitis G. Resolving the limitation of the peak fitting and peak shape methods in the determination of the activation energy of thermoluminescence glow peaks. *J. Lumin.* 2014b; (146): 418-423.
- [20] Sadek A M, Eissa H M, Basha A M, Kitis G. Properties of the thermoluminescence glow peaks simulated by the interactive multiple-trap system (IMTS) model. *Phys. Stat. Sol. (B)*. 2014a; (252): 721-729.
- [21] Sadek A M, Khamis F, Polymeris G S, Carinou E, Kitis G. Similarities and differences between two different types of the thermoluminescence dosimeters belonging to the LiF family. *Phys. Stat. Sol. (C)*. 2016; 14(1-2):1-10 / DOI 101002/PSSC.201600220.
- [22] Sadek A M, Kitis G. A critical look at the kinetic parameter values used in simulating the thermoluminescence glow-curve. *J. Lumin.* 2017b; (183): 533-541.
- [23] Sadek A M, Hassan M M, Esmat E, Eissa H M. A new approach to the analysis of thermoluminescence glow-curve of TLD-600 dosimeters following Am-241 alpha particles irradiation. *Radiat. Prot. Dosim.* 2017a; 178 (3): 260-271.
- [24] Balian H G, Eddy N W. Figure-of-merit (FOM), an improved criterion over the normalized Chi-squared test for assessing goodness-of-fit of Gamma-ray spectral peaks. *Nucl. Instrum and Meth in Phys. Res.* 1977; (145): 389-395.
- [25] Horowitz Y S, Yossian D. Computerized glow curve deconvolution: Application to thermoluminescence dosimetry. *Radiat. Prot. Dosim.* 1995; (60): 1-114.
- [26] Yossian D, Horowitz Y S. Computerized glow curve deconvolution applied to the analysis of the kinetics of peak 5 in LiF: Mg, Ti. *J. Phys. D: Appl. Phys.* 1995; (28): 1495-1508.
- [27] Bakshi A K, Pradhan A S, Srivastava K, Kolambe D H. New parameters of annealing and re-use of TLDs based on CaSO₄: Dy phosphor. *Radiat. Prot. Dosim.* 2002; (100): 293-296.
- [28] Bos A J J, Piters T M, Gómez-Ros J M, Delgado A. An intercomparison of glow curve analysis computer programs: II. Measured glow curves. *Radiat. Prot. Dosim.* 1994; (51):257-264.
- [29] Perks C A, Marshall M. Techniques for thermoluminescence glow curve analysis. *Radiat. Prot. Dosim.* 1991; (38): 261-269.
- [30] Horowitz Y S, Satinger D, Yossian D, Brandan M E, Buenfill A E, Gamboa-deBuen I and et al. Ionization density effects in the thermoluminescence of TLD-100: computerized T_m-T_{stop} glow curve analysis. *Radiat. Prot. Dosim.* 1999; (48): 239-242.
- [31] Horowitz Y S, Satinger D, Fuks E, Oster L, Podpalov L. On the use of LiF: Mg, Ti thermoluminescence doseimeters in space - A critical review. *Radiat. Prot. Dosim.* 2003; (106): 7-24.
- [32] Yossian D, Mahajna S, Shachar B.B, Horowitz Y S. Re-investigation of the kinetic trapping parameters of peak 5 in TLD-100 via "prompt" and "residual" isothermal decay. *Radiat. Prot. Dosim.* 1993; (47): 129-133.
- [33] Sadek A M, Eissa H M, Basha A M, Carinou E, Askounis P, Kitis G. The deconvolution of thermoluminescence glow-curves using general expressions derived from the one trap-one recombination (OTOR) level model. *Appl. Radiat. Isot.* 2015; (95): 214-221.
- [34] Sadek A M, Kitis G. Impact of non-fulfillment of the superposition principle on the analysis of thermoluminescence glow-curve. *Radiat. Meas.* 2018; (116): 14-23.

المخلص العربي

مها فرج¹، عمر محمد صادق¹، هاني عبدالله شوشه¹، عزة عبد الروؤف²، جورج كيتيس³

يعتبر قياس الجرعات الأشعاعية عن طريق كواشف الوميض الحرارى باستخدام مادة الليثيوم فلورايد المطعم بالماغنسيوم والتيتانيوم من افضل الطرق فى تقييم وقياس الجرعة الأشعاعية الممتصة للعاملين بالأشعة المؤينة .

لذلك لوحظ إجراء العديد من الأبحاث لتحسين وفهم خصائص تلك المواد المستخدمة فى قياس الجرعات الأشعاعية وكذلك لتطوير مواد جديدة. وقد وجد ان مقاييس الجرعات الإشعاعية باستخدام كواشف الوميض الحرارى عند تعريضها للأشعاعات عالية الجرعة نليها المعالجة الحرارية يمكن تعزيزها كأحدى الطرق فى تحسين خصائص تلك المواد وتسمى هذه الطريقة بزيادة حساسية المواد للإشعاع ؛ وهناك طرق عديدة اخرى تم تطبيقها لتعزيز وزيادة حساسية مقاييس الجرعة باستخدام (TLD-100)

وقد قام (11) بتطوير إجراء لتحسين حساسية مقياس الجرعة لمادة (TLD-100) من خلال تعريضها لجرعة إشعاعية عالية مقدارها (2 كيلو جراى) ويليها معالجة حرارية عند درجة حرارة عالية مقدارها 553 كلفن لمدة (30 دقيقة) وهذا الإجراء أدى إلى زيادة فى عامل الحساسية (S/S₀) للمساحة الكلية تحت منحنى التوهج بمقدار 35 ضعف .

وأیضا زيادة لعامل الحساسية للقيمة السابعة التى تقع فى مدى نطاق القيم العالية لدرجات الحرارة لمنحنى التوهج الحرارى بمقدار 22.

والهدف الأساسى لإجراء زيادة الحساسية لمقاييس TLD بهذه الطريقة هو تقليل تأثير المصايد العميقة بالألكترونات التى قد تكون بمثابة المنافس للمصايد المنخفضة عند التشعيع بجرعات عالية مما يساهم فى ملء المصايد النشطة لذلك عند التشعيع اللاحق يتم ملء المصايد النشطة دون التدخل فى المنافسة مع المصايد العميقة وهذا الإجراء يزيد من الحساسية ويعمل على توسيع منطقة النطاق الخطى وبالتالي إذا كان التشعيع الأول للعينات يتسبب فى الأضرار بالمصايد العميقة المتاحة مما يؤدي إلى زيادة عدد الشحنات التى يتم حبسها لكل وحدة تعريض فى المصايد النشطة وهو ما يهدف إليه مما يؤدي إلى زيادة فى حساسية العينات.

ومن المعروف ان مقاييس الجرعة يتم إعادة استخدامها مرات عديدة لنفس المواد عن طريق تطبيق معالجة حرارية لتفريغها من جميع الألكترونات الناتج من التعريض للجرعات الأشعاعية ولكن من أضرار تلك الأستخدام المتكرر مع تكرار المعالجة الحرارية عند كل استخدام جديد أنه يبيؤثر بشكل كبير وملحوظ على حساسية تلك المقاييس مع الزمن.

لذلك يقوم تلك البحث بدراسة موثوقية تأثير الأستخدام المتكرر عدة مرات مع اختلاف مدى النطاق الحرارى لقراءة اشارات الوميض الحرارى لمادة TLD-100 المعالجة بالفعل بإستخدام طريقة تحسين الحساسية التى تم ذكرها من قبل حيث تتم القراءة باستخدام جهاز الوميض الحرارى عند النطاق الحرارى (323-553).

وذلك مع زيادة تلك النطاق كل مرة بمقدار 20 درجة حرارية حتى تصل إلى 633 كلفن مع تكرار تلك العملية عشر مرات. لذلك كان الهدف الرئيسى من دراسة تلك البحث هو دراسة فحص إجراء المعالجة السابقة لمادة الليثيوم فلورايد التى تم تطبيقها فى [11].

من خلال شقين:

- 1- فحص تأثير القراءة لإشارات الوميض الحرارى مع اختلاف مدى النطاق الحرارى المطبق.
- 2- فحص تأثير الإستخدام المتكرر على عامل الحساسية والعوامل الحركية وذلك على كل قمة من قمم منحى التوهج الحرارى وأيضا على المساحة الكلية تحت المنحنى.

وتضمنت النتائج أن فى حالة القراءة عند نطاق درجات الحرارة (323-553) كلفن أن القمة الخامسة هى القمة السائدة والمهيمنة والأعلى قيمة فى منحى التوهج الحرارى ولكن فى حالة نطاق القراءة عند درجات الحرارة (323-633) كلفن تصبح منطقة الحرارة العالية جدا لمنحنى التوهج الحرارى هى المنطقة السائدة والأعلى قيمة للمنحنى.

وقد تم الإستنتاج انه بعد إعادة الأستخدام 10 مرات تقل حساسية المساحة الكلية تحت منحى التوهج الحرارى لمادة الليثيوم فلورايد المعالجة بمقدار 0.6 وذلك لجميع حالات القراءة لإشارات الوميض الحرارى عند نطاق درجات الحرارة المختلفة ماعدا حساسية منطقة القراءة لإشارات الوميض الحرارى عند درجات الحرارة للمدى (323-633) كلفن لا تتأثر بعدد مرات الأستخدام.

وأظهرت النتائج أيضا أن اشارات الوميض الحرارى للمنطقة الحرارية المنخفضة تقل ولكن تزيد مع المنطقة الحرارية العالية جدا وذلك نتيجة تأثير إعادة الأستخدام مقاييس الجرعات عدة مرات.

وأثبتت النتائج أيضا أن المساحة الكلية تحت منحى التوهج الحرارى تصبح مستقرة ومستقلة أيضا بعد حوالى 7 مرات من إعادة الأستخدام ولا تتأثر قيمتها أيضا بعد هذا الرقم من تكرار الأستخدام.

وقد بينت النتائج أيضا أنه فى حالة إعادة إستخدام مقاييس الليثيوم فلورايد المعالجة قد تؤثر على حساسيتها حتى فى ظل نفس الظروف التجريبية وهذا عندما لا تشمل القراءة الكلية لإشارات الوميض الحرارى نطاق قراءة المنطقة الحرارية العالية لمنحنى التوهج الحرارى وذلك عندما تكون درجة الحرارة النهائية للقراءة أكبر من 505 كلفن ، لكن إذا تضمنت القراءة لإشارات الوميض الحرارى لتلك المنطقة ذات درجة الحرارة العالية فوجد أن إشارات الوميض الحرارى لمنطقة الحرارة العالية جدا تزيد بشكل كبير مع زيادة عدد إعادة إستخدام مقاييس الجرعات حتى تصل إلى 7 مرات تصبح تلك المنطقة مستقرة وموثوقة تقريبا للأستخدام فى قياس الجرعات الإشعاعية.

لذلك يوصى عند القراءة بأن تكون درجة الحرارة النهائية للقياس تقترب او اكبر من 633 كلفن لتشمل منطقة إشارات الوميض الحرارى لدرجات الحرارة العالية لمقاييس الليثيوم فلورايد المعالجة حيث ان تلك المنطقة تظل ثابتة مع الأستخدام المتكرر عدة مرات.

وقد وجد أيضا أن إعادة الأستخدام والقراءة بمدى مختلف من درجات الحرارة لنفس المقاييس ليس له أى تأثير على العوامل الحركية لمادة الليثيوم فلورايد المعالجة ، ولكن لوحظ أيضا زيادة قيم طاقة التنشيط للقمة الخامسة من قمم منحى التوهج الحرارى وهذه الزيادة ليست بسبب اختلاف مدى نطاق درجات الحرارة لقراءة إشارات الوميض الحرارى ولا بسبب عدد مرات إعادة الأستخدام ولكن ترجع الزيادة إلى عملية المعالجة المطبقة على مادة الليثيوم فلورايد والتي تم ذكرها من قبل.

